

**June 16, 2023**

## **QUESTIONS FOR AIR MONITORING EQUIPMENT, SUPPLIES & SERVICES**

- In your RFP, the first piece of equipment you are requesting states *"An ambient air monitoring package that includes measurements using a FEM ozone monitor, plus CO<sub>2</sub>, CO, PM<sub>1</sub>/PM<sub>2.5</sub>, and total VOC sensors that fit inside a ventilated, mountable outdoor enclosure. The power requirement should be 15 watts or less. There should be an option to upload data to the Cloud via cellular access. Figure 1 RFP: For Air Monitoring Equipment, Supplies, and Services Page 12 Pricing should include the option for cellular communication service and data access for the duration of the project."* **Question:** Since EPA FEM designation for Ozone requires instrumentation to be deployed in an air conditioned enclosure, will you accept bid responses that include all of the requirements except FEM designation for the Ozone monitor? (If an FEM Ozone monitor/analyzer is deployed in a non-conforming/non-air conditioned enclosure, it will not meet FEM designation). We have a strong solution and would like to submit a bid response, however we would like some advice given this odd requirement.

**Yes, we will accept bids that include an ozone monitor without the FEM designation.**

- On Page 12 of the RFP you have preferred sensor specifications for Carbon Monoxide sensors. Under "Precision" you listed "Contact 2B Tech for information." **Question:** Was this a typo?

**Yes, this was a typo. I used information from several vendor websites to build the specifications and this was a copy/past error. Please disregard the precision requirement.**